



हर कदम, हर डगर
किसानों का हमसफर
भारतीय कृषि अनुसंधान परिषद
AgriSearch with a human touch

BY SPEED POST/FAX

भारतीय मृदा विज्ञान संस्थान (भा0कृ0अनु0प0)
नबीबाग, बैरसिया रोड, भोपाल – 462038

Indian Institute of Soil Science(ICAR)
Nabibagh, Berasia Road, Bhopal – 462038 (M.P.)

Tel. No.(0755)2747375 EPABX:2730970/2734221 (Ext. No. 233 & 261) Fax. No. (0755) 2733310

F.No.5-497/2017-18/P&S/IISS

Dated: 16.05.2018

To,

Sub: - Quotation for purchase of **pH Meter with Redox Electrode.** – reg.

Dear Sir(s),

Reputed firms/Authorized firms are invited for quoting their rates for the supply of following stores as per under mentioned specifications: -

Name & Specification of the Item	Qty.	Rate Rs.	Total Rs.																																						
<p>pH Meter with Redox Electrode.</p> <p>Features</p> <ul style="list-style-type: none"> ✓ Microprocessor Based ✓ Operating key pad ✓ mV measurement range-2000 mV ✓ Up to 3-point calibration with auto-buffer recognition ✓ Compliance Certifications ✓ The pH electrode should be for soil analysis. 	01																																								
<table border="1"> <thead> <tr> <th colspan="2">Specifications:</th> </tr> <tr> <th>Product specification</th> <th>Description</th> </tr> </thead> <tbody> <tr> <td>pH Range</td> <td>0 - 14</td> </tr> <tr> <td>Resolution</td> <td>0.01 pH</td> </tr> <tr> <td>Accuracy</td> <td>±0.05</td> </tr> <tr> <td>Calibration</td> <td>Up to 3 points with auto buffer</td> </tr> <tr> <td>Modes</td> <td>pH mV-C</td> </tr> <tr> <td colspan="2">Temperature</td> </tr> <tr> <td>Operating temperature</td> <td>0-40 °C</td> </tr> <tr> <td>Temperature Accuracy</td> <td>≤ 0.3 °C</td> </tr> <tr> <td>Temperature compensation Range</td> <td>Automatic (0 to 100 Degree c)</td> </tr> <tr> <td>Range (meter)</td> <td>0.0 to 100.0 °C/32.0 to 212. °F</td> </tr> <tr> <td>Operating Temperature Degree C</td> <td>45</td> </tr> <tr> <td>Compensation</td> <td>ATC/MTC</td> </tr> <tr> <td>Display</td> <td>LED/LCD/Touch Screen, 3 Digit</td> </tr> <tr> <td>Impact Resistant</td> <td>Yes</td> </tr> <tr> <td>Test Reports for EMC/EMI As Per EN 61326-1:2013</td> <td>Yes</td> </tr> <tr> <td>Test Report No and date</td> <td>W6M21102-11252-E11</td> </tr> <tr> <td>Power requirements</td> <td>230 V ± 10, 50 Hz AC</td> </tr> </tbody> </table>				Specifications:		Product specification	Description	pH Range	0 - 14	Resolution	0.01 pH	Accuracy	±0.05	Calibration	Up to 3 points with auto buffer	Modes	pH mV-C	Temperature		Operating temperature	0-40 °C	Temperature Accuracy	≤ 0.3 °C	Temperature compensation Range	Automatic (0 to 100 Degree c)	Range (meter)	0.0 to 100.0 °C/32.0 to 212. °F	Operating Temperature Degree C	45	Compensation	ATC/MTC	Display	LED/LCD/Touch Screen, 3 Digit	Impact Resistant	Yes	Test Reports for EMC/EMI As Per EN 61326-1:2013	Yes	Test Report No and date	W6M21102-11252-E11	Power requirements	230 V ± 10, 50 Hz AC
Specifications:																																									
Product specification	Description																																								
pH Range	0 - 14																																								
Resolution	0.01 pH																																								
Accuracy	±0.05																																								
Calibration	Up to 3 points with auto buffer																																								
Modes	pH mV-C																																								
Temperature																																									
Operating temperature	0-40 °C																																								
Temperature Accuracy	≤ 0.3 °C																																								
Temperature compensation Range	Automatic (0 to 100 Degree c)																																								
Range (meter)	0.0 to 100.0 °C/32.0 to 212. °F																																								
Operating Temperature Degree C	45																																								
Compensation	ATC/MTC																																								
Display	LED/LCD/Touch Screen, 3 Digit																																								
Impact Resistant	Yes																																								
Test Reports for EMC/EMI As Per EN 61326-1:2013	Yes																																								
Test Report No and date	W6M21102-11252-E11																																								
Power requirements	230 V ± 10, 50 Hz AC																																								

Contd...

Terms & conditions are as below: -

01. The rates should be quoted F.O.R. IISS, Nabi Bagh, Bhopal basis along with indicating separately brand name its price, packing forwarding, freight, tax percentage and insurance etc. in clear terms.
02. Quantity of discount, if any, should be indicated clearly.
03. This institute is registered with the Department of Scientific & Industrial Research (DSIR), Govt. of India for purposes of availing custom duty exemption in terms of Government Notification No 51/96-Customs dated 23 July 1996 and central excise duty exemption in terms of Government Notification No. 10/97-Central Excise dated 1 March 1997.
04. Rates of taxes/GST etc. Should be indicated in clear terms separately. GST no. and other relevant registration number both under State and Central Tax Act/Rules should be separately indicated.
05. **Taxes:** As per notification No. 45/2017-Central Tax(Rate) dated 14.11.2017 issued by the Min. of Finance, Scientific and Technical instruments, apparatus, equipment (including computers) are exempted from the Central tax (in excess of 2.5%) leviable there on under section 9 of the Central Goods and Service Tax Act,2017 supplied to the Departments & Laboratories of the Central Government and State Governments, other than a hospital. Since, this Institute is a Scientific Research Institution under Indian Council of Agricultural Research, Min. of Agriculture Govt. of India, and also registered by Min. of Science and Technology, DSIR, New Delhi vide their Registration Certificate No. TU/V/RG-CDE (504)/2016 dated 09.01.2017 **the GST has to be charged accordingly.**
06. Unless otherwise mentioned in the quotation, it will be presumed that the quoted rates are inclusive of all taxes/levies and free delivery/installation at our Institute.
07. The rates quoted should be valid for 120 days from the date of opening of quotation.
08. There should be no cutting/ overwriting. The cutting if any should be duly attested.
09. Quotation should be submitted along with printed literature/ catalogue of the quoted make/ model and authorization certificate of the manufacturer.
10. Unattested/amended/overwritten figures will not be considered.
11. In all matters of disputes, the decision of the Director of this Institute shall be final and binding.
12. Minimum period of delivery of the stores/materials should clearly be specified in the quotation.
13. TDS etc. will be deducted from the bill as per applicable tax rules of GOI/M.P. Govt.
14. **PAYMENT TERMS:** Payment will be made within reasonable time say within 30 days from the date of receipt of material in good condition without any damage and after installation of the equipment in accordance with the ordered specification at the end. **Since the mode of payment will be e-payment system, hence Bank Account Number, name of bank with location and IFSC Code Number of Bank may be furnished on the body of the bill.**
15. Quotation should be sent by in a sealed cover and the covering envelope should contain "Quotation for **pH Meter with Redox Electrode** vide letter **F.No.5-497/2017-18/P&S/IISS due on 15.06.2018**. Quotation received in an unsealed cover or received after the due date and those, which are ambiguous, are liable to be rejected. **The last date of receipt of above quotation is 15.06.2018**. This institute will not be responsible for any postal/speed post/courier delays.
16. Bid Security (EMD)-in the shape of DD worth Rs. 2,000/- (Two Thousand Only) **in favour of "ICAR UNIT-IISS Bhopal"** must be submitted along with the quotation otherwise the bid will not be taken into consideration.
17. Performance Security: Within 10 (Ten) days after the issue of a letter of award by this Institute, the supplier, shall furnish performance security @ 10 % of ordered value in the form of Demand Draft in favour of "ICAR UNIT-IISS Bhopal" including the warranty obligation.
18. The Institute reserves the right to reject any or all the quotation(s) received without assigning any reason thereof.

Yours Sincerely,

Senior Administrative Officer

Copy to: -

1. I/c AKMU for uploading on IISS, website & CPP Portal.
2. Nodal Officer, CPP Portal, IISS, Bhopal for information and needful.